

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00167	Calib. Date: 03/12/2010
Certification File No.:	00168	Cert. No.: 00008
Linearity File No.:	00169	Lin. Date: 03/12/2010
Solution File No.:	00175	Soln. No.: 00075
Sequential File No.:	00175	File Date: 06/20/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0283
Solution Control Lot:	09L072	Expires: 12/31/2011
		Bottle No.: 0395

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:01D		
Control 1 EC	0.097%	18:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:02D		
Control 2 EC	0.097%	18:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		
Control 3 EC	0.098%	18:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FALCICCHIO

First Name: NICHOLAS

MI: -

Signature: _____

Nicholas Falcicchio
06/20/2010

Badge No.: 45

Date: 06/20/2010