

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWE-0027 |
| Location: | BLAIRSTOWN POLICE DEPT. | |
| Calibration File No.: | 00178 | Calib. No.: 00011 |
| Certification File No.: | 00179 | Cert. No.: 00009 |
| Linearity File No.: | 00180 | Lin. No.: 00009 |
| Solution File No.: | 00184 | Soln. No.: 00078 |
| Sequential File No.: | 00184 | File Date: 08/29/2010 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWH S3-0284 |
| Solution Control Lot: | 09L072 | Expires: 12/31/2011 |
| | | Bottle No.: 0396 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 20:16D | | |
| Control 1 EC | 0.100% | 20:16D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 20:16D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:17D | | |
| Control 2 EC | 0.100% | 20:17D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 20:17D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:18D | | |
| Control 3 EC | 0.100% | 20:19D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 20:19D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:19D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FALCICCHIO

First Name: NICHOLAS

MI: -

Signature: _____



Badge No.: 45

Date: 08/29/2010