

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWE-0027
Location: BLAIRSTOWN POLICE DEPT.
Calibration File No.: 00178 Calib. Date: 08/05/2010 Calib. No.: 00011
Certification File No.: 00179 Cert. Date: 08/05/2010 Cert. No.: 00009
Linearity File No.: 00180 Lin. Date: 08/05/2010 Lin. No.: 00009
Solution File No.: 00191 Soln. Date: 12/04/2010 Soln. No.: 00082
Sequential File No.: 00191 File Date: 12/04/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0284
Control Solution %: 0.100% Expires: 12/31/2011
Solution Control Lot: 09L072 Bottle No.: 0393

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:12S		
Control 1 EC	0.100%	21:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:13S		
Control 2 EC	0.100%	21:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15S		
Control 3 EC	0.100%	21:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FALCICCHIO

First Name: NICHOLAS

MI: -

Signature: _____



Badge No.: 45

Date: 12/04/2010