

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWE-0027  
 Location: BLAIRSTOWN POLICE DEPT.  
 Calibration File No.: 00110 Calib. Date: 10/30/2008 Calib. No.: 00007  
 Certification File No.: 00111 Cert. Date: 10/30/2008 Cert. No.: 00005  
 Linearity File No.: 00112 Lin. Date: 10/30/2008 Lin. No.: 00005  
 Solution File No.: 00128 Soln. Date: 03/18/2009 Soln. No.: 00055  
 Sequential File No.: 00128 File Date: 03/18/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0283  
 Control Solution %: 0.100% Expires: 03/07/2010  
 Solution Control Lot: 08C051 Bottle No.: 0954

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:16S		
Control 1 EC	0.098%	16:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18S		
Control 2 EC	0.097%	16:18S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19S		
Control 3 EC	0.098%	16:20S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DALRYMPLE

First Name: KYLE

MI: B.

Signature: \_\_\_\_\_

Badge No.: 43

Date: 03/18/2009