

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00129	Calib. Date: 04/03/2009
Certification File No.:	00130	Calib. No.: 00008
Linearity File No.:	00131	Cert. Date: 04/03/2009
Solution File No.:	00135	Cert. No.: 00006
Sequential File No.:	00135	Lin. Date: 04/03/2009
		Lin. No.: 00006
		Soln. Date: 05/14/2009
		Soln. No.: 00059
		File Date: 05/14/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0284
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0969

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:13D		
Control 1 EC	0.099%	14:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15D		
Control 2 EC	0.099%	14:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:16D		
Control 3 EC	0.099%	14:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DALRYMPLE

First Name: KYLE

MI: B.

Signature: _____

Badge No.: 43

Date: 05/14/2009