

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWE-0027
Location: BLAIRSTOWN POLICE DEPT.
Calibration File No.: 00129 Calib. Date: 04/03/2009 Calib. No.: 00008
Certification File No.: 00130 Cert. Date: 04/03/2009 Cert. No.: 00006
Linearity File No.: 00131 Lin. Date: 04/03/2009 Lin. No.: 00006
Solution File No.: 00137 Soln. Date: 06/09/2009 Soln. No.: 00060
Sequential File No.: 00137 File Date: 06/09/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0284
Control Solution %: 0.100% Expires: 03/07/2010
Solution Control Lot: 08C051 Bottle No.: 0970

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:21D		
Control 1 EC	0.099%	14:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:22D		
Control 2 EC	0.098%	14:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:24D		
Control 3 EC	0.099%	14:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DALRYMPLE

First Name: KYLE

MI: B.

Signature: _____



Badge No.: 43

Date: 06/09/2009