

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.		
Calibration File No.:	00167	Calib. Date:	03/12/2010
Certification File No.:	00168	Cert. Date:	03/12/2010
Linearity File No.:	00169	Lin. Date:	03/12/2010
Solution File No.:	00177	Soln. Date:	07/29/2010
Sequential File No.:	00177	File Date:	07/29/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0283
Solution Control Lot:	09L072	Expires:	12/31/2011
		Bottle No.:	0376

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:06D		
Control 1 EC	0.098%	15:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 2 EC	0.097%	15:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08D		
Control 3 EC	0.098%	15:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FALCICCHIO

First Name: NICHOLAS

MI: -

Signature: _____



Badge No.: 45

Date: 07/29/2010