

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.		
Calibration File No.:	00148	Calib. Date:	09/28/2009
Certification File No.:	00149	Cert. Date:	09/28/2009
Linearity File No.:	00150	Lin. Date:	09/28/2009
Solution File No.:	00158	Soln. Date:	01/15/2010
Sequential File No.:	00158	File Date:	01/15/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWH S3-0283
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0968

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:48S		
Control 1 EC	0.102%	10:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49S		
Control 2 EC	0.100%	10:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51S		
Control 3 EC	0.099%	10:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DALRYMPLE

First Name: KYLE

MI: B

Signature: _____

Badge No.: 43

Date: 01/15/2010