

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00110	Calib. Date: 10/30/2008
Certification File No.:	00111	Calib. No.: 00007
Linearity File No.:	00112	Cert. Date: 10/30/2008
Solution File No.:	00123	Cert. No.: 00005
Sequential File No.:	00123	Lin. Date: 10/30/2008
		Lin. No.: 00005
		Soln. Date: 01/23/2009
		Soln. No.: 00053
		File Date: 01/23/2009

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0283
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0965

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:03S		
Control 1 EC	0.100%	16:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04S		
Control 2 EC	0.100%	16:05S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05S		
Control 3 EC	0.100%	16:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LOSEY

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_



Badge No.: 36

Date: 01/23/2009