

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00148	Calib. Date: 09/28/2009
Certification File No.:	00149	Calib. No.: 00009
Linearity File No.:	00150	Cert. Date: 09/28/2009
Solution File No.:	00153	Cert. No.: 00007
Sequential File No.:	00153	Lin. Date: 09/28/2009
		Lin. No.: 00007
		Soln. Date: 10/27/2009
		Soln. No.: 00066
		File Date: 10/27/2009

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWH S3-0283
Control Solution %: 0.100%		Expires: 03/07/2010
Solution Control Lot: 08C051		Bottle No.: 0975

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:49S		
Control 1 EC	0.100%	18:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:51S		
Control 2 EC	0.100%	18:51S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:51S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:52S		
Control 3 EC	0.101%	18:53S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:53S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DALRYMPLE

First Name: KYLE

MI: B.

Signature: _____

Badge No.: 43

Date: 10/27/2009