

Calibrating Unit New Standard Solution Report

Serial No.: ARWE-0027

Calib. No.: 00007

Cert. No.: 00005

Lin. No.: 00005

Soln. No.: 00051

Calib. Date: 10/30/2008

Cert. Date: 10/30/2008

Lin. Date: 10/30/2008

Soln. Date: 11/24/2008

File Date: 11/24/2008

Model No.: CU-34

Serial No.: DWH S3-0283

Expires: 03/07/2010

Bottle No.: 0957

Equipment

Alcoteq 7110 MKIII-C

Location: BLAIRSTOWN POLICE DEPT.

Calibration File No.: 00110

Certification File No.: 00111

Linearity File No.: 00112

Solution File No.: 00118

Sequential File No.: 00118

Calibrating Unit: WET

Control Solution %: 0.100%

Solution Control Lot: 08C051

Function	Result	Time	Temperature	Comment(s)
Ambient Air Blank	0.000%	13:36S		
Control 1 EC	0.099%	13:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38S		
Control 2 EC	0.099%	13:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39S		
Control 3 EC	0.099%	13:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcoteq 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSEY



First Name: STEPHEN

Badge No.: 36

Date: 11/24/2008

MI: P