

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00110	Calib. Date: 10/30/2008
Certification File No.:	00111	Calib. No.: 00007
Linearity File No.:	00112	Cert. Date: 10/30/2008
Solution File No.:	00121	Cert. No.: 00005
Sequential File No.:	00121	Lin. Date: 10/30/2008
		Lin. No.: 00005
		Soln. Date: 12/27/2008
		Soln. No.: 00052
		File Date: 12/27/2008

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWH S3-0283
Control Solution %:	0.100%		Expires: 03/07/2010
Solution Control Lot:	08C051		Bottle No.: 0964

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:08S		
Control 1 EC	0.100%	11:08S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		
Control 2 EC	0.099%	11:10S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10S		
Control 3 EC	0.099%	11:11S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LOSEY

First Name: STEPHEN

MI: P

Signature: \_\_\_\_\_

Badge No.: 36

Date: 12/27/2008