

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWE-0027
Location: BLAIRSTOWN POLICE DEPT.
Calibration File No.: 00110 Calib. Date: 10/30/2008 Calib. No.: 00007
Certification File No.: 00111 Cert. Date: 10/30/2008 Cert. No.: 00005
Linearity File No.: 00112 Lin. Date: 10/30/2008 Lin. No.: 00005
Solution File No.: 00125 Soln. Date: 02/20/2009 Soln. No.: 00054
Sequential File No.: 00125 File Date: 02/20/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWH S3-0283
Control Solution %: 0.100% Expires: 03/07/2010
Solution Control Lot: 08C051 Bottle No.: 0960

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:59S		
Control 1 EC	0.099%	12:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00S		
Control 2 EC	0.097%	13:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02S		
Control 3 EC	0.099%	13:02S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSEY

First Name: STEPHEN

MI: P

Signature: 

Badge No.: 36

Date: 02/20/2009