

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00148	Calib. No.: 00009
Certification File No.:	00149	Cert. No.: 00007
Linearity File No.:	00150	Lin. No.: 00007
Solution File No.:	00163	Soln. No.: 00070
Sequential File No.:	00163	File Date: 02/03/2010

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0283
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0952

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:19S		
Control 1 EC	0.100%	18:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:20S		
Control 2 EC	0.100%	18:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:22S		
Control 3 EC	0.100%	18:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DALRYMPLE

First Name: KYLE

MI: B

Signature: _____

Badge No.: 43

Date: 02/03/2010