

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00148	Calib. Date: 09/28/2009
Certification File No.:	00149	Cert. No.: 00009
Linearity File No.:	00150	Cert. Date: 09/28/2009
Solution File No.:	00166	Lin. Date: 09/28/2009
Sequential File No.:	00166	Lin. No.: 00007
		Soln. No.: 00071
		File Date: 03/06/2010

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0283
Solution Control Lot:	09L072	Expires: 12/31/2011
		Bottle No.: 0397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:09S		
Control 1 EC	0.100%	23:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:10S		
Control 2 EC	0.098%	23:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:12S		
Control 3 EC	0.100%	23:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FALCICCHIO

First Name: NICHOLAS

MI: -

Signature: _____



Badge No.: 45

Date: 03/06/2010