

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWE-0027
Location:	BLAIRSTOWN POLICE DEPT.	
Calibration File No.:	00129	Calib. No.: 00008
Certification File No.:	00130	Cert. No.: 00006
Linearity File No.:	00131	Lin. No.: 00006
Solution File No.:	00140	Soln. No.: 00062
Sequential File No.:	00140	File Date: 08/04/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWH S3-0284
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0961

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:05D		
Control 1 EC	0.098%	17:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:06D		
Control 2 EC	0.099%	17:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07D		
Control 3 EC	0.098%	17:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DALRYMPLE

First Name: KYLE

MI: B.

Signature: \_\_\_\_\_

Badge No.: 43

Date: 08/04/2009